

**From:** Cynthia Caporale/ESC/R3/USEPA/US  
**Sent:** 4/25/2012 5:51:47 PM  
**To:** Richard Fetzer/R3/USEPA/US@EPA  
**CC:**  
**Subject:** Fw: Dimock Round 2 sampling

Here is another email about the micro.

Also, we wanted to clarify that phosphorus was total and not ortho-phosphorus.

Cindy

Cynthia Caporale, Chief  
OASQA Laboratory Branch  
U.S. EPA Region III  
Environmental Science Center  
Fort Meade, MD  
(410) 305-2732  
Fax: (410) 305-3095

----- Forwarded by Cynthia Caporale/ESC/R3/USEPA/US on 04/25/2012 05:50 PM -----

**From:** Dave Russell/ESC/R3/USEPA/US  
**To:** Cynthia Caporale/ESC/R3/USEPA/US@EPA  
**Cc:** Annie Hilliard/ESC/R3/USEPA/US  
**Date:** 04/25/2012 12:56 PM  
**Subject:** Re: Fw: Dimock Round 2 sampling

Maybe I missed it but how many samples are involved for bacteria? Couldn't find a number.

There would be a 30 hour holding time whether total coliform rule or the ground water rule is applied.

Need to add E.coli to analytes determined. Colisure determines p/a for both total coliform and e. coli.